



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7G00AZ

TECHNOTHROMBIN TGA RD Reagenz 5x1.5mL

REF

5006220

LOT

7G51C00.01



2027-03-31

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA RD Reagenz	7G51C00	2027-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z53C00	410 - 683	553	OK
	TGA Control Low (nM Peak Thrombin) LOT 2852C00	404 - 673	496	OK

Durchführung von
Prepared by

Tikret NASUFI *Tikret Nasufi*

08. JULI 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ *Lucas*

08. JULI 2025

Datum / Date: