

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7G00AZ
---	---	----------


TECHNOTHROMBIN TGA RD Reagenz 5x1.5mL

REF 5006220
LOT 7G51C00.01
 2027-03-31

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA RD Reagenz	7G51C00	2027-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z53C00	410 - 683	553	OK
	TGA Control Low (nM Peak Thrombin) LOT 2852C00	404 - 673	496	OK

Durchführung von
Prepared by


Mehmet NASUFI

08. JULI 2025
Datum / Date:

Freigabe von
Approved by


Lucas HOLOVICZ

08. JULI 2025
Datum / Date: