

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN® TGA SUB 5x1.5 mL

REF 5006235
LOT 8F41C00.02
 2026-04-30

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA Substrat	8F41C00	2026-04-30

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z42C00	TGA RC High LOT 7F52C00	340 - 566	465	OK
	TGA Control Low (nM Peak Thrombin) LOT 2842C00	TGA RC High LOT 7F52C00	170 - 284	223	OK

Durchführung von
Prepared by

Fikret NASUFI 

04. FEB. 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOKOVICZ 

04. FEB. 2025

Datum / Date: