



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8F00AZ

TECHNOTHROMBIN TGA SUB 5x1.5 mL

REF

5006235

LOT

8F51C00.01



2027-02-28

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA Substrat	8F51C00	2027-02-28

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z42C00	TGA RC Low LOT 7D31C01	250 - 416	378	OK
		TGA RC High LOT 7F52C00	340 - 566	447	OK
	TGA Control Low (nM Peak Thrombin) LOT 2842C00	TGA RC Low LOT 7D31C01	119 - 198	148	OK
		TGA RC High LOT 7F52C00	170 - 284	226	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

06. MAI 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOLWICZ *Lucas Holwicz*

06. MAI 2025

Datum / Date: