



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8F00AZ

TECHNOTHROMBIN® TGA SUB 5x1.5 mL

REF 5006235

LOT 8F41C00.01

🕒 2026-04-30

Inhalt Content	Reagenz Reagent	LOT	🕒
5x1.5 mL	TGA Substrat	8F41C00	2026-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z41C00	TGA RC Low LOT 7D31C01	269 - 448	417.0	OK
		TGA RC High LOT 7F21C01	332 - 553	512.0	OK
	TGA Control Low (nM Peak Thrombin) LOT 2841C00	TGA RC Low LOT 7D31C01	94 - 214	188.8	OK
		TGA RC High LOT 7F21C01	151 - 308	233.2	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

14. JUNI 2024

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ *Lucas HoloVICZ*

14. JUNI 2024

Datum / Date: