



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1Z00AZ

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320

LOT 1Z42C00.01

2025-08-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z42C00	2025-08-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC High (nM Peak Thrombin) LOT 7F31C00	340 - 566	504 OK
	TGA RC Low (nM Peak Thrombin) LOT 7D31C01	250 - 416	319 OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

10. OKT. 2024

Datum / Date:

Freigabe von  
Approved by

Maylis RAUSCHER *Maylis Rauscher*

10. OKT. 2024

Datum / Date: