

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Z00AZ
--	---	----------


TECHNOTHROMBIN TGA Control High 5x1mL

REF 5006320
LOT 1Z53C00.01
 2026-05-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z53C00	2026-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC High (nM Peak Thrombin) LOT 7F53C00	422 - 704	633	OK
	TGA RC Low (nM Peak Thrombin) LOT 7D51C00	312 - 520	465	OK

Durchführung von
Prepared by

Fikret NASUFI 

08. JULI 2025

Datum / Date:

Freigabe von
Approved by


LUCAS HOLOVICZ

08. JULI 2025

Datum / Date: