
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Z00AZ
---	---	----------

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320
LOT 1Z41C00.01
 2025-02-28

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z41C00	2025-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC High (nM Peak Thrombin) LOT 7F21C01	332 - 553	512.0	OK
	TGA RC Low (nM Peak Thrombin) LOT 7D31C01	269 - 448	417.0	OK

Durchführung von
Prepared by

Ejket NASUFI 

14. JUNI 2024
Datum / Date:

Freigabe von
Approved by

Maylis RAUSCHER 

14. JUNI 2024
Datum / Date: