technoclone TC	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Z00AZ

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320

LOT 1Z42C00.01

⌛ 2025-08-31

Inhalt Content	Reagenz Reagent	LOT	⌛
5x1mL	TGA Control High	1Z42C00	2025-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC High (nM Peak Thrombin) LOT 7F31C00	340 - 566	504	OK
	TGA RC Low (nM Peak Thrombin) LOT 7D31C01	250 - 416	319	OK

Durchführung von
Prepared by

Fikret NASUFI

10. OKT. 2024
Datum / Date:

Freigabe von
Approved by

Marlis RAUSCHER

10. OKT. 2024
Datum / Date: