



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2800AZ

TECHNOTHROMBIN® TGA Control Low 5x1mL

REF 5006330

LOT 2842C00.01

🕒 2025-08-31

| Inhalt Content | Reagenz Reagent | LOT | 🕒 |
|-------------------|--------------------|---------|------------|
| 5x1mL | TGA Control Low | 2842C00 | 2025-08-31 |

| Analytische Prozedur Analytical Procedure | Anforderung Requirement | Ergebnis Result | | |
|---|--|--------------------|-----|----|
| Kontrollwiederfindung Recovery of controls | TGA RC High (nM Peak Thrombin) LOT 7F31C00 | 170 - 284 | 251 | OK |
| | TGA RC Low (nM Peak Thrombin) LOT 7D31C01 | 119 - 198 | 191 | OK |

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

10. OKT. 2024

Datum / Date:

Freigabe von
Approved by

Maylis RAUSCHER *Maylis Rauscher*

10. OKT. 2024

Datum / Date: