

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2800AZ
---	---	----------

TECHNOTHROMBIN® TGA Control Low 5x1mL

REF 5006330
LOT 2852C00.01
 2026-04-30

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control Low	2852C00	2026-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC High (nM Peak Thrombin) LOT 7F53C00	240 - 416	368	OK
	TGA RC Low (nM Peak Thrombin) LOT 7D51C00	121 - 299	220	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

08. JULI 2025
Datum / Date:

Freigabe von
Approved by

Lucas
LUCAS HOLOVICZ

08. JULI 2025
Datum / Date: