

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0D00AZ
---	---	----------

TECHNOTHROMBIN® TGA CAL Set

REF 5006345
LOT 0D42C00.01
 2025-11-30

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA CAL BUF	9P41C00	2026-01-31
1x0.5mL	TGA CAL	5T31C	2025-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kalibrierung / Calibration gegen / against WHO 19/188		report	960	OK	
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197074	Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	1.00	OK	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z41C00	TGA RCL Reagent LOT 7D31C01	269 - 448	383.3	OK
		TGA RCH Reagent LOT 7F21C01	332 - 553	488.5	OK
	TGA Control Low (nM Peak Thrombin) LOT 2841C00	TGA RCL Reagent LOT 7D31C01	94 - 214	170.1	OK
		TGA RCH Reagent LOT 7F21C01	151 - 308	272.5	OK


Durchführung von
Prepared by

Fikret NASUFI 

14. JUNI 2024

Datum / Date:

Freigabe von
Approved by



14. JUNI 2024

Datum / Date: