



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0D00AZ

TECHNOTHROMBIN® TGA CAL Set

REF 5006345
LOT 0D43C00.01
🕒 2026-01-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x3mL	TGA CAL BUF	9P41C00	2026-01-31
1x0.5mL	TGA CAL	5T41C	2026-06-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kalibrierung / Calibration gegen / against WHO 19/188		report	877	OK	
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	1.00	OK	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z42C00	TGA RCL Reagent LOT 7D31C01	250 - 416	296.6	OK
		TGA RCH Reagent LOT 7F31C00	340 - 566	467.8	OK
	TGA Control Low (nM Peak Thrombin) LOT 2842C00	TGA RCL Reagent LOT 7D31C01	119 - 198	145.3	OK
		TGA RCH Reagent LOT 7F31C00	170 - 284	232.9	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

10. OKT. 2024

Datum / Date:

Freigabe von
Approved by

Maylis BAUSCHER *Maylis Bauscher*

10. OKT. 2024

Datum / Date: