

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0D00AZ
---	---	----------

TECHNOTHROMBIN TGA CAL Set

REF 5006345
LOT 0D51C00.01
 2026-06-30

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA CAL BUF	9P51C00	2027-03-31
1x0.5mL	TGA CAL	5T41C	2026-06-30

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 19/188			report	877	OK
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve		≥ 0.99	1.00	OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z42C00	TGA RCL Reagent LOT 7D31C01	250 - 416	310	OK
		TGA RCH Reagent LOT 7F52C00	340 - 566	421	OK
	TGA Control Low (nM Peak Thrombin) LOT 2842C00	TGA RCL Reagent LOT 7D31C01	119 - 198	148	OK
		TGA RCH Reagent LOT 7F52C00	170 - 284	196	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

19. SEP. 2025

Datum / Date:

Freigabe von
Approved by

Lucas *Lucas* MOLOVICZ

19. SEP. 2025

Datum / Date: