

CERTIFICATE OF ANALYSIS

Technothrombin TGA CAL Set



Attachment 25 DQ0T00.02

Temp.: TF0327.07

0D61C00.02

REF	5006345
LOT	0D61C00.02
	2027-03-31

Content	volume	Reagent	LOT	
1x	3 mL	Technothrombin TGA Cal BUF	9P51C00	2027-03-31
1x	0.5 mL	Technothrombin TGA Cal	5T51C	2027-11-30

Analytical Procedures	Requirement	Result
visual inspection	Content	according to packaging SOP confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Vial barcode	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Box barcode	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Batch table	present confirmed OK
	Recovery of controls	Technothrombin TGA Control High Lot: 1Z61C00 461 - 769 nM
Technothrombin TGA RCH	Technothrombin TGA Control Low Lot: 2861C00 222 - 384 nM	351 nM OK

Prepared by Fikret NASUFI *Fikret Nasufi* 17. APR. 2026
Date

Approved by Maylis RAUSCHER *Maylis Rauscher* 17. APR. 2026
Date