

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0700AZ
---	---	----------

Ceveron TGA CAL Set

REF 5006347-RUO

LOT 0752C00.02

 2026-12-31

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	Ceveron TGA CAL BUF	9S41C00	2026-12-31
1x0.5mL	Ceveron TGA CAL	5Y41C	2026-12-31

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration Gegen / against WHO 19/188			report	864	OK
Verwendbarkeit / usability Gerät / Analyzer: Ceveron 100 series	Regression der Bezugskurve / regression of the calibration curve		≥ 0.990	1.0	OK
Kontrollwiederfindung Recovery of controls [LOT] 1Y54C00	Peak Thrombin nM	Ceveron® TGA RC High (nM Peak Thrombin) [LOT] 7Q41C	396 - 660	509	OK
Kontrollwiederfindung Recovery of controls [LOT] 2752C00		Ceveron® TGA RC High (nM Peak Thrombin) [LOT] 7Q41C	214 - 384	279	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

16. JUNI 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOLVICH

16. JUNI 2025

Datum / Date: