

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9P00AZ
---	---	----------

TECHNOTHROMBIN TGA BUF 5x3 mL

REF 5006360
LOT 9P41C00.02
2026-01-31

Inhalt Content	Reagenz Reagent	LOT	2026-01-31
5 x 3 mL	TGA Buffer	9P41C00	2026-01-31

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197091		Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	1.00	OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z42C00	TGA RC High LOT 7F52C00	340 - 566	487	OK
	TGA Control Low (nM Peak Thrombin) LOT 2842C00	TGA RC High LOT 7F52C00	170 - 284	241	OK

Durchführung von
Prepared by

Erret NASUFI *Erret Nasufi*

15. MAI 2025

Datum / Date:

Freigabe von
Approved by

Maylis RAUSCHER *Maylis Rauscher*

16. MAI 2025

Datum / Date: