

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9P00AZ
---	---	----------

## TECHNOTHROMBIN® TGA BUF 5x3 mL

REF      5006360  
LOT        9P21C01.01  
      2025-04-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5 x 3 mL	TGA Buffer	9P21C01	2025-04-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result			
Verwendbarkeit / usability  Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve	$\geq 0.99$	0.99	OK	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z24C00	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F21C00	407 - 679	459.4	OK
	TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2823C00	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F21C00	127 - 211	139.4	OK


**Durchführung von  
Prepared by**

Michael KAFKA 

20. APR. 2023

Datum / Date:

**Freigabe von  
Approved by**

Fikret NASUFI 

20. APR. 2023

Datum / Date: