
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9P00AZ
---	---	----------


## TECHNOTHROMBIN® TGA BUF 5x3 mL

REF      5006360  
LOT        9P41C00.01  
      2026-01-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5 x 3 mL	TGA Buffer	9P41C00	2026-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result			
Verwendbarkeit / usability  Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	1.00	OK	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z42C00	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F31C00	340 - 566	512	OK
	TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2842C00	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F31C00	170 - 284	266	OK


**Durchführung von  
Prepared by**

Fikret NASUFI 

10. OKT. 2024

Datum / Date:

**Freigabe von  
Approved by**

Maylis RAUSCHER 

10. OKT. 2024

Datum / Date: