

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9P00AZ
---	---	----------

TECHNOTHROMBIN TGA BUF 5x3 mL

REF 5006360
LOT 9P51C00.01
 2027-03-31

Inhalt Content	Reagenz Reagent	LOT	
5 x 3 mL	TGA Buffer	9P51C00	2027-03-31

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197074		Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	1.00	OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z53C00	TGA RC High LOT 7F53C00	422 - 704	618	OK
	TGA Control Low (nM Peak Thrombin) LOT 2852C00	TGA RC High LOT 7F53C00	240 - 416	358	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

28. OKT. 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ *Lucas HoloVICZ*

28. OKT. 2025

Datum / Date: