CERTIFICATE OF ANALYSIS
Technothrombin TGA BUF

technoclone

Attachment 26 DQ0T00.02

Temp.: TF0326.07

9P51C00.02

REF	5006360
LOT	9P51C00.02
⌛	2027-03-31

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	white	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	clear liquid	confirmed	OK
Recovery of controls Technothrombin TGA RCH	**Technothrombin TGA Control High** **Lot:** 1Z54C00	536 - 894 nM	638 nM	OK
	Technothrombin TGA Control Low **Lot:** 2853C00	246 - 410 nM	341 nM	OK

Prepared by Fikret NASUFI Fikret Nasufi 03. APR. 2026
Date

Approved by Lukas GRUBELNIG 03. APR. 2026
Date

Approved by