

# CERTIFICATE OF ANALYSIS

## AK-Calibrant



Attachment 07 DQ0U00.05

0U51C00.03

Temp.: YF0327 07

REF 5010004-RUO  
 LOT 0U51C00.03  
 2026-11-30

Content	volume	Reagent	LOT	
1x	1 mL	AK-Calibrant Level A	1N42C00	2026-11-30
1x	1 mL	AK-Calibrant Level B	3L46C00	2026-11-30
1x	1 mL	AK-Calibrant Level C	3M45C00	2026-11-30
1x	1 mL	AK-Calibrant Level D	3N46C00	2026-12-31

Analytical Procedures		Requirement		Result	
visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
Recovery of controls TECHNOPLASTIN HIS	<b>Control N</b> LOT 1P31C01	0.83	- 1.11 INR	0.96 INR	OK
	<b>Control A</b> LOT 3P42C00	2.02	- 2.72 INR	2.46 INR	OK

Prepared by

Stefan BRUCKBERGER

26. MRZ. 2025

Date

Approved by

Michael KAFKA

26. MRZ. 2025

Date