


CERTIFICATE OF ANALYSIS
AK-Calibrant




Attachment 07 DQ0U00.06

Temp.: TF0327.07

0U54C00.03

REF 5010004-RUO
 LOT 0U54C00.03
 2027-07-31

Content	volume	Reagent	LOT	
1x	1 mL	AK-Calibrant Level A	1N42C01	2027-07-31
1x	1 mL	AK-Calibrant Level B	3L46C01	2027-07-31
1x	1 mL	AK-Calibrant Level C	3M45C01	2027-07-31
1x	1 mL	AK-Calibrant Level D	3N46C01	2027-07-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Recovery of controls TECHNOPLASTIN HIS	Control N		
LOT 1A54C00		0.88 - 1.16 INR	1.04 INR	OK
Control A				
	LOT 3A51C00	2.21 - 2.95 INR	2.64 INR	OK

Prepared by


Lukas GRUBELNIG

20. FEB. 2026

Date

Approved by


Marcin MANUGIEWICZ

20. FEB. 2026

Date