CERTIFICATE OF ANALYSIS Coagulation Control N 5x1 mL



Attachment 04

DQ1P00.03

1P31C01.03

REF

5020040-RUO

LOT

1P31C01.03

 \mathbf{Z}

2026-06-30

Analytical Procedures		Requirement	Result	
	Cap color	red	confirmed	ОК
visual inspection	Vial type	clear glass vial	confirmed	ОК
	Vial label	according to packaging SOP	confirmed	ОК
	Vial barcode	according to packaging SOP	confirmed	ОК
	Box label	according to packaging SOP	confirmed	ОК
	Box barcode	according to packaging SOP	confirmed	ОК
	Package insert	present	confirmed	ОК
	Batch table	present	confirmed	ОК
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	ок
	Appearance after reconstitution	amber colored liquid	confirmed	ОК
	HBsAg / HBsAg	Negative	Negative	
Virology	HIV-Ak / HIV-Ab	Negative	Negative	
	HCV-AK / HCV-Ab	Negative	Negative	ОК
The testing methods a	pplied were FDA-approved or CE marked.			
Recovery of controls	Siron LS	20.1 - 27.3 sec	22.3 sec	ОК

Prepared by	Maylis RAUSCHER	0 2. OKT. 2024
,	/ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \	Date
Approved by	LIVEST COVEZ	0 Z. UKT. 202 4
and the street of the street		Date