## CERTIFICATE OF ANALYSIS Coagulation Control N 5x1 mL



Attachment 04

DQ1P00.03

1P31C01.02

REF

5020040-RUO

LOT

1P31C01.02

23

2026-06-30

| Analytical Procedures  |                                       | Requirement                                  | Result      | Result |  |
|------------------------|---------------------------------------|--|-------------|--------|--|
| visual inspection      | Cap color                             | red  | confirmed   | ОК     |  |
|                        | Vial type                             | clear glass vial                             | confirmed   | ОК     |  |
|                        | Vial label                            | according to packaging SC                    | P confirmed | ОК     |  |
|                        | Vial barcode                          | according to packaging SC                    | P confirmed | ОК     |  |
|                        | Box label                             | according to packaging SC                    | P confirmed | ОК     |  |
|                        | Box barcode                           | according to packaging SC                    | P confirmed | ОК     |  |
|                        | Package insert                        | present                                      | confirmed   | ОК     |  |
|                        | Batch table                           | present                                      | confirmed   | ОК     |  |
|                        | Appearance before reconstitution      | buff colored plug of lyophilized<br>material | confirmed   | ок     |  |
|                        | Appearance after reconstitution       | amber colored liquid                         | confirmed   | ОК     |  |
|                        | HBsAg / HBsAg                         | Negative                                     | Negative    |        |  |
| Virology               | HIV-Ak / HIV-Ab                       | Negative                                     | Negative    |        |  |
|                        | HCV-AK / HCV-Ab                       | Negative                                     | Negative    | ОК     |  |
| The testing methods ap | plied were FDA-approved or CE marked. |  |             |        |  |
| Recovery of controls   | Siron LS                              | 20.1 - 27.3 se                               | ec 24.1 sec | ОК     |  |

| Prepared by  | Stefan BRUCKBERGER | 2 6. JUNI 2024 |
|--|--------------------|----------------|
| an ing ang managan ang man |                    | Date           |
| Approved by  | Maylis RAUSCHER    | 2 7. JUNI 2024 |
|  |                    | Date           |