

CERTIFICATE OF ANALYSIS

Coagulation Control A



Attachment 04 DQ3P00.06

3P51C00.02

Temp.: T103256.07

REF 5021055-RUO
LOT 3P51C00.02
2027-04-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology	HBsAg / HBsAg	Negative	Negative	OK
	HIV-Ak / HIV-Ab	Negative	Negative	
	HCV-AK / HCV-Ab	Negative	Negative	
The testing methods applied were FDA-approved or CE marked.				
Recovery of controls	Siron LS	31.8 - 43.0 sec	40.7 sec	OK

Prepared by

Lavdri NASUFI

18. AUG. 2025

Date

Approved by

Lukas GRUBELNIG

19. AUG. 2025

Date