CERTIFICATE OF ANALYSIS
Coagulation Control A 5x1 ml

technoclone

Attachment 04 DQ3P00.05

Temp. : TF0326.07

3P42C00.02

REF	5021055-RUO
LOT	3P42C00.02
⧖	2026-02-28

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology The testing methods applied were FDA-approved or CE marked.	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Recovery of controls	**Siron LS**	37.5 - 50.7 sec	43.8 sec	OK

Prepared by Maylis RAUSCHER 18. APR. 2024
Date

Approved by Michael KAFKA 18. APR. 2024
Date