



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ6V0A.05

Att. 05

6V32CA0.03

Dapttin TC 5 x 2 mL

REF 5035060-RUO
LOT 6V32CA0.03
⌚ 2025-07-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Cap color	grey	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	turbid liquid	confirmed	OK
Kontrollwiederfindung aPTT am Ceveron [®] alpha	Control N LOT 1P11C01	25.8 - 35.0 sec	32.6 sec	OK
Recovery of controls aPTT on Ceveron [®] alpha	Control A LOT 3P23C00	59.7 - 80.7 sec	72.8 sec	OK

Durchgeführt von
Prepared by

Maylis RAUSCHER

18. APR. 2024

Datum / Date:

Freigabe von
Approved by

Smilja DURANOVIĆ

18. APR. 2024

Datum / Date: