

Technoview Arixtra CAL



Attachment 16 DQ2600.02

2642C00.02

Temp.: TF0327.06

REF 5090010-RUO
LOT 2642C00.02
2026-04-30

Content	volume	Reagent	LOT	
1x	1mL	Technoview Arixtra CAL 1	2141C00	2026-04-30
1x	1mL	Technoview Arixtra CAL 2	2241C00	2026-04-30
1x	1mL	Technoview Arixtra CAL 3	2341C00	2026-04-30
1x	1mL	Technoview Arixtra CAL 4	2441C00	2026-04-30
1x	1mL	Technoview Arixtra CAL 5	2541C00	2026-04-30

Analytical Procedures		Requirement		Result	
visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
Recovery of controls	Arixtra CON H LOT 3V21C01	0.98	- 1.64 µg/mL	1.46 µg/mL	OK
	Arixtra CON L LOT 3U21C01	0.34	- 0.64 µg/mL	0.42 µg/mL	OK

Prepared by

Maylis RAUSCHER

18. JUNI 2024

Date

Approved by

Lucas HOLOVIEZ

18. JUNI 2024

Date