

Technoview Arixtra CAL



Attachment 16 DQ2600.02

2651C00.01

Temp.: TF0327.06

REF 5090010-RUO
 LOT 2651C00.01
 2027-06-30

Content	volume	Reagent	LOT	
1x	1mL	Technoview Arixtra CAL 1	2141C01	2027-06-30
1x	1mL	Technoview Arixtra CAL 2	2241C01	2027-06-30
1x	1mL	Technoview Arixtra CAL 3	2341C01	2027-06-30
1x	1mL	Technoview Arixtra CAL 4	2441C01	2027-06-30
1x	1mL	Technoview Arixtra CAL 5	2541C01	2027-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	Arixtra CON H LOT 3V41C01	0.97 - 1.61 µg/mL	1.32 µg/mL	OK
	Arixtra CON L LOT 3U41C01	0.35 - 0.65 µg/mL	0.45 µg/mL	OK

Prepared by

Maylis BAUSCHER

14. AUG. 2025

Date

Approved by

Lucas FLOVICZ

14. AUG. 2025

Date