## **Technoview Arixtra CON L**



mp.: TF0326.06

3U41C01.01

REF

5090012-RUO

LOT

3U41C01.01

2

2027-06-30

Analytical Procedures		Requirement	Result	
	Cap color	red	confirmed	ОК
visual inspection	Vial type	clear glass vial	confirmed	ОК
	Vial label	according to packaging SOP	confirmed	ОК
	Vial barcode	according to packaging SOP	confirmed	ОК
	Box label	according to packaging SOP	confirmed	ОК
	Box barcode	according to packaging SOP	confirmed	ОК
	Package insert	present	confirmed	ОК
	Batch table	present	confirmed	ОК
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	ОК
	Appearance after reconstitution	amber colored liquid	confirmed	ОК
	HBsAg / HBsAg	Negative	Negative	
Virology	HIV-Ak / HIV-Ab	Negative	Negative	
	HCV-AK / HCV-Ab	Negative	Negative	ОК
The testing methods a	applied were FDA-approved or CE marked.			
Recovery of control		0.35 - 0.65 μg/mL	0.51 µg/ml	ОК

Prepared by	Mayis RAUSCHER	3 O. JUL1 2025
(100 AS 4) 4 (4) € (20 AS 5) (100 Aug 46) a vertic € (4)		Date
Approved by	Lucas HOLOVICZ	3 1. JULI <b>202</b> 5
	<u> </u>	Date