

CERTIFICATE OF ANALYSIS
Technoview LMWH CAL



Attachment 14 DQ0H00.02

Temp.: TFG327.07

0H41C00.03

REF	5090040-RUO
LOT	0H41C00.03
⌚	2026-04-30

Content	volume	Reagent	LOT	⌚
1x	1 mL	Technoview LMWH CAL 1	5A41C00	2026-04-30
1x	1 mL	Technoview LMWH CAL 2	5B41C00	2026-04-30
1x	1 mL	Technoview LMWH CAL 3	5C41C00	2026-04-30
1x	1 mL	Technoview LMWH CAL 4	5D41C00	2026-04-30
1x	1 mL	Technoview LMWH CAL 5	5E41C00	2026-04-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	LMWH CON H LOT 3H41C00	0.83 - 1.39 IU/mL	1.11 IU/mL	OK
	LMWH CON L LOT 3F41C00	0.34 - 0.64 IU/mL	0.43 IU/mL	OK

Prepared by Stefan BRUCKBERGER

27. MRZ. 2025
Date

Approved by Fikret NASUFI

27. MRZ. 2025
Date