

CERTIFICATE OF ANALYSIS
Technoview LMWH CAL



Attachment 14 DQ0H00.02
0H51C00.01

Temp.: TPO227.07

REF 5090040-RUO
LOT 0H51C00.01
⌚ 2027-06-30

Content	volume	Reagent	LOT	⌚
1x	1 mL	Technoview LMWH CAL 1	5A41C01	2027-06-30
1x	1 mL	Technoview LMWH CAL 2	5B41C01	2027-06-30
1x	1 mL	Technoview LMWH CAL 3	5C41C01	2027-06-30
1x	1 mL	Technoview LMWH CAL 4	5D41C01	2027-06-30
1x	1 mL	Technoview LMWH CAL 5	5E41C01	2027-06-30

Analytical Procedures		Requirement		Result	
visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
Recovery of controls	LMWH CON H LOT 3H32C01	0.77	- 1.28 IU/mL	1.01 IU/mL	OK
	LMWH CON L LOT 3F32C02	0.25	- 0.47 IU/mL	0.33 IU/mL	OK

Prepared by Maylis RAUSCHER

30. JULI 2025
Date

Approved by Lucas HOLOVICZ

30. JULI 2025
Date