

# CERTIFICATE OF ANALYSIS

## Technoview LMWH CON L



Attachment 17 DQ0H00.02

3F41C00.02

Temp.: TFO326.07

REF 5090042-RUO  
 LOT 3F41C00.02  
 2026-04-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology		HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	OK
The testing methods applied were FDA-approved or CE marked.				
Recovery of control		0.34 - 0.64 IU/mL	0.55 IU/mL	OK

Prepared by

*Maya AUSCHER*

08. MAI 2025

Date

Approved by

*Lucas HOLOWICZ*

08. MAI 2025

Date