
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3C00.02
---	---	-----------


### TECHNOVIEW UFH Control L 5x1 mL

REF      5090072-RUO  
LOT      3C21C01.02  
      2025-07-31

<b>Analytische Prozedur</b> <b>Analytical Procedure</b>		<b>Anforderung</b> <b>Requirement</b>	<b>Ergebnis</b> <b>Result</b>	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® Alpha	Anti-Xa UFH	0.21 – 0.39 IU/mL	0.33	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

Michael KAFKA



03. AUG. 2023

Datum / Date:

**Freigabe von**  
**Approved by**

  
Suzana GRESAK PODAKAR

03. AUG. 2023

Datum / Date: