CERTIFICATE OF ANALYSIS
Technoview UFH CON L

technoclone

Attachment 13 DQ0Y00.02

Temp.: TF0326.07

3C41C00.03

REF 5090072-RUO

LOT 3C41C00.03

2026-02-28

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology	HBsAg / HBsAg	Negative	Negative	OK
	HIV-Ak / HIV-Ab	Negative	Negative	
	HCV-AK / HCV-Ab	Negative	Negative	
The testing methods applied were FDA-approved or CE marked.				
Recovery of control		0.13 - 0.25 IU/mL	0.15 IU/mL	OK

Prepared by Maylis RAUSCHER 07. NOV. 2024 Date

Approved by Lucas HOLOVICZ 07. NOV. 2024 Date