

**CERTIFICATE OF ANALYSIS**  
**Technoview UFH CON H**



Attachment 12      DQ0Y00.02

Temp.: TFD026.07

3J42C00.02

REF	5090074-RUO
LOT	3J42C00.02
⌚	2026-02-28

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
Appearance after reconstitution	amber colored liquid	confirmed	OK	
Virology	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab  The testing methods applied were FDA-approved or CE marked.	Negative Negative Negative	Negative Negative Negative	OK
Recovery of control	0.41 - 0.68 IU/mL	0.52 IU/mL	OK	

Prepared by Maylis RAUSCHER

02. OKT. 2024  
Date

Approved by Lucas HOŁOWICZ

02. OKT. 2024  
Date