

CERTIFICATE OF ANALYSIS
Technoview UFH CON H




Attachment 12 DQ0Y00.02

Temp.: TFO226.C7


3J42C00.03

REF	5090074-RUO
LOT	3J42C00.03
⌚	2026-02-28

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
The testing methods applied were FDA-approved or CE marked.				
Recovery of control	0.41 - 0.68 IU/mL	0.57 IU/mL	OK	

Prepared by  Malis RAUSCHER

07. NOV. 2024
Date

Approved by  Lukas HOFMEIER

07. NOV. 2024
Date