

	<h1>ANALYSENZERTIFIKAT</h1> <h2>CERTIFICATE OF ANALYSIS</h2>	0241C00.02
---	--	------------

Technoview® Argatroban CAL Set 5x1mL

REF	5090140-RUO
LOT	0241C00.02
⌚	2026-04-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	⌚
			0241C00	
1x	1mL	Argatroban CAL 1	5Q11C02	2026-04-30
1x	1mL	Argatroban CAL 2	5R11C02	2026-04-30
1x	1mL	Argatroban CAL 3	5S11C02	2026-04-30
1x	1mL	Argatroban CAL 4	5U11C02	2026-04-30
1x	1mL	Argatroban CAL 5	5V11C02	2026-04-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Kontrollwiederfindung am Ceveron® Alpha Recovery of controls on Ceveron® Alpha	Technoview® Argatroban CONT L LOT 3111C02	0.43 - 0.79 µg/mL	0.62 µg/mL	OK
	Technoview® Argatroban CONT H LOT 3211C02	0.86 - 1.44 µg/mL	1.19 µg/mL	OK

Durchgeführt von
Prepared by



22. APR. 2024

Datum / Date:

Freigabe von
Approved by



22. APR. 2024

Datum / Date: