

CERTIFICATE OF ANALYSIS

Technoview Rivaroxaban CAL



Attachment 17

DQ0300.03

Temp.: TR0326.07

REF 5090170-RUO
 LOT 0951C00.02
 2027-06-30

Content	volume	Reagent	LOT	
1x	1mL	Technoview Rivaroxaban CAL 1	5141C01	2027-06-30
1x	1mL	Technoview Rivaroxaban CAL 2	5641C01	2027-06-30
1x	1mL	Technoview Rivaroxaban CAL 3	5741C01	2027-06-30
1x	1mL	Technoview Rivaroxaban CAL 4	5841C01	2027-06-30
1x	1mL	Technoview Rivaroxaban CAL 5	5941C01	2027-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	Rivaroxaban CON M LOT 3531C02	87.8 - 146.4 ng/mL	123.7 ng/mL	OK
	Rivaroxaban CON L LOT 3451C00	44.4 - 82.4 ng/mL	65.7 ng/mL	OK

Prepared by

Maylis RAUSCHER

20. OKT. 2025

Date

Approved by

Fikret NASUFI

20. OKT. 2025

Date