

CERTIFICATE OF ANALYSIS
Technoview Rivaroxaban CAL H



Attachment 19 DQ0300.03

Temp.: TFO326.07

REF 5090171-RUO
LOT 0341C00.02
⌚ 2026-06-30

Content	volume	Reagent	LOT	⌚
1x	1mL	Technoview Rivaroxaban high CAL 1	5131C01	2026-06-30
1x	1mL	Technoview Rivaroxaban high CAL 2	5231C01	2026-06-30
1x	1mL	Technoview Rivaroxaban high CAL 3	5331C01	2026-06-30
1x	1mL	Technoview Rivaroxaban high CAL 4	5431C01	2026-06-30
1x	1mL	Technoview Rivaroxaban high CAL 5	5531C01	2026-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	Rivaroxaban CON H LOT 3321C02	226.1 - 376.8 ng/mL	296.0 ng/mL	OK
	Rivaroxaban CON M LOT 3511C01	78.1 - 130.1 ng/mL	97.9 ng/mL	OK

Prepared by Lucas HOLOVIAZ

31. OCT. 2024
Date

Approved by Stefan BRUCKBERGER

04. NOV. 2024
Date