

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3400.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control Low 5x1 mL

REF 5090172-RUO

LOT 3421C01.02

 2025-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifischer Kalibrationswert Lot specific calibration value		40.1 – 74.5 ng/mL	57.3	OK
Verwendbarkeit Usability Ceveron 100 series	TECHNOCHROM® anti-Xa	40.1 – 74.5 ng/mL	59.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Suzana GRESAK PODAKAR

31. AUG. 2023

Datum / Date:

Freigabe von
Approved by


LUCAS HOLOWICZ

31. AUG. 2023

Datum / Date: