

CERTIFICATE OF ANALYSIS

Technoview Rivaroxaban CON H



Temp.: 170326.07

Attachment 22 DQ0300.03

3321C02.03

REF 5090174-RUO
 LOT 3321C02.03
 2026-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab The testing methods applied were FDA-approved or CE marked.		Negative Negative Negative	Negative Negative Negative	OK
Recovery of control		226.1 - 376.8 ng/mL	313.4 ng/mL	OK

Prepared by

Lucas HOLOVICZ

08. OKT. 2025

Date

Approved by

Maylis RAUSCHER

06. OKT. 2025

Date