
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control High 5x1 mL

REF 5090174-RUO
LOT 3321C01.02
 2025-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		226.1 – 376.8 ng/mL	301.4	
Verwendbarkeit Usability Ceveron 100 series	TECHNOVIEW® anti-Xa	226.1 – 376.8 ng/mL	310.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Suzana GRESAK PODAKAR

31. AUG. 2023

Datum / Date:

Freigabe von
Approved by


Lucas HOLYICZ

31. AUG. 2023

Datum / Date: