

CERTIFICATE OF ANALYSIS

Technoview Dabigatran CAL 4x1 mL



Attachment 11 DQ0600.01

Temp.: FF0326.07

0641C00.02

REF	5090210-RUO
LOT	0641C00.02
	2026-03-31

Content	volume	Reagent	LOT	
1x	1mL	Technoview Dabigatran CAL 1	1T22C02	2026-03-31
1x	1mL	Technoview Dabigatran CAL 2	1U22C02	2026-03-31
1x	1mL	Technoview Dabigatran CAL 3	1V22C02	2026-03-31
1x	1mL	Technoview Dabigatran CAL 4	1W22C02	2026-03-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	Dabigatran CON H LOT 3622C01	289.5 - 482.5 ng/mL	401.4 ng/mL	OK
	Dabigatran CON L LOT 3722C01	86.8 - 161.2 ng/mL	121.7 ng/mL	OK

Prepared by

Lucas HOLOWICZ

10. APR. 2024

Date

Approved by

Lavdriq NASUFI

10. APR. 2024

Date