

# CERTIFICATE OF ANALYSIS

## Technoview Apixaban CAL



Attachment 11

DQ1000.01

Temp.: YF0326.07

REF 5090269-RUO  
 LOT 1051C00.02  
 2027-06-30

| Content | volume | Reagent                   | LOT     |            |
|---------|--------|---------------------------|---------|------------|
| 1x      | 1mL    | Technoview Apixaban CAL 1 | 1141C01 | 2027-06-30 |
| 1x      | 1mL    | Technoview Apixaban CAL 2 | 1241C01 | 2027-06-30 |
| 1x      | 1mL    | Technoview Apixaban CAL 3 | 1341C01 | 2027-06-30 |
| 1x      | 1mL    | Technoview Apixaban CAL 4 | 1441C01 | 2027-06-30 |
| 1x      | 1mL    | Technoview Apixaban CAL 5 | 1541C01 | 2027-06-30 |

| Analytical Procedures |                                       | Requirement                | Result       |    |
|-----------------------|---------------------------------------|----------------------------|--------------|----|
| visual inspection     | Content                               | according to packaging SOP | confirmed    | OK |
|                       | Vial label                            | according to packaging SOP | confirmed    | OK |
|                       | Vial barcode                          | according to packaging SOP | confirmed    | OK |
|                       | Box label                             | according to packaging SOP | confirmed    | OK |
|                       | Box barcode                           | according to packaging SOP | confirmed    | OK |
|                       | Package insert                        | present                    | confirmed    | OK |
|                       | Batch table                           | present                    | confirmed    | OK |
| Recovery of controls  | <b>Apixaban CONT H</b><br>LOT 3D41C01 | 290.10 - 483.50 ng/mL      | 436.80 ng/mL | OK |
|                       | <b>Apixaban CONT L</b><br>LOT 3941C01 | 110.30 - 204.90 ng/mL      | 153.40 ng/mL | OK |

Prepared by

Maylis RAUSCHER

29. JULI 2025

Date

Approved by

Lucas HOLOVICZ

29. JULI 2025

Date