

CERTIFICATE OF ANALYSIS

Technoview Apixaban CAL



Attachment 11

DQ1000.01

Temp.: YF0326.07

REF 5090269-RUO
LOT 1051C00.02
2027-06-30

Content	volume	Reagent	LOT	2027-06-30
1x	1mL	Technoview Apixaban CAL 1	1141C01	2027-06-30
1x	1mL	Technoview Apixaban CAL 2	1241C01	2027-06-30
1x	1mL	Technoview Apixaban CAL 3	1341C01	2027-06-30
1x	1mL	Technoview Apixaban CAL 4	1441C01	2027-06-30
1x	1mL	Technoview Apixaban CAL 5	1541C01	2027-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	Apixaban CONT H LOT 3D41C01	290.10 - 483.50 ng/mL	436.80 ng/mL	OK
	Apixaban CONT L LOT 3941C01	110.30 - 204.90 ng/mL	153.40 ng/mL	OK

Prepared by

Maylis RAUSCHER

29. JULI 2025

Date

Approved by

Lucas HOLOVICZ

29. JULI 2025

Date