

CERTIFICATE OF ANALYSIS

Technoview Apixaban CON H



Attachment 12 DQ1000.01

3D21C01.02

Temp.: T0325.07

REF 5090270-RUO
LOT 3D21C01.02
2026-02-28

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology		HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	negative negative negative
The testing methods applied were FDA-approved or CE marked.				OK
Recovery of control		219.4 - 365.6 ng/mL	276.0 ng/mL	OK

Prepared by

Lucas HOLWICZ

04. MRZ. 2024

Date

Approved by

Michael KAFKA

04. MRZ. 2024

Date