

# CERTIFICATE OF ANALYSIS

## Technoview Apixaban CON H



Attachment 12 DQ1000.01

Temp.: TF0326.07

3D41C01.02

REF 5090270-RUO  
 LOT 3D41C01.02  
 2027-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
The testing methods applied were FDA-approved or CE marked.				
Recovery of control		290.1 - 483.5 ng/mL	436.8 ng/mL	OK

Prepared by

Maylis RAUSCHER

29. JULI 2025

Date

Approved by

Lucas HOLOVICZ

29. JULI 2025

Date