CERTIFICATE OF ANALYSIS
FP: Fibrinogen Reagent Kit

technoclone

Attachment 10 DQ6E0A.03

Temp.: TF0227.07

0B42C00.02

REF	5138005-RUO
LOT	0B42C00.02
⌛	2025-12-31

Content	volume	Reagent	LOT	⌛
5x	2 mL	Fibrinogen Reagent	6E36CA0	2025-12-31
1x	1 mL	Coagulation Reference	1R45C00	2026-08-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	**Control N** LOT 1P31C01	2.67 - 4.01 g/L	3.36 g/L	OK
	Control A LOT 3P42C00	0.93 - 1.39 g/L	1.22 g/L	OK

Prepared by Maylis RAUSCHER 27. SEP. 2024 Date

Approved by Lucas HOLOVAC 27. SEP. 2024 Date